



NOTES

1. MATERIAL: UVFS
2. CLEAR APERTURE: 22 mm
3. CENTER APERTURE: 365 nm
4. FWHM: 10 nm
5. TRANSMITTED WAVEFRONT: $\lambda @ 633$ nm
6. SURFACE QUALITY: 60/40(S/D)
7. COATING: BANDPASS FILTER, 0° AOI
 Tabs $\geq 90\% @ 365$ nm;
 ODavg $\geq 5 @ 300$ nm–355 nm;
 ODabs $\geq 5, ODavg \geq 6 @ 375$ nm–800 nm;

DRAWING PROJECTION			LBTEK			
	NAME	DATE	ET365/10x			
DRAWN	LZHOU	JUL./1st/24	PREMIUM BANDPASS FILTER $\phi 25$ mm, CWL=365 nm, FWHM=10 nm			
APPROVAL	WCHENG	JUL./1st/24	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			UVFS	1.5 g	2:1	A