



NOTES

1. MATERIAL: UVFS
2. CLEAR APERTURE: 22 mm
3. CENTER APERTURE: 514 nm
4. FWHM: 10 nm
5. TRANSMITTED WAVEFRONT: $\lambda @ 633$ nm
6. SURFACE QUALITY: 60/40(S/D)
7. COATING: BANDPASS FILTER, 0° AOI
 Tabs $\geq 95\% @ 514$ nm;
 ODabs ≥ 5 , ODavg $\geq 6 @ 300$ nm–501 nm & 527 nm–900 nm;

DRAWING PROJECTION			LBTEK			
	NAME	DATE				
DRAWN	LZHOU	JUL./1st/24	PREMIUM BANDPASS FILTER $\phi 25$ mm, CWL=514 nm, FWHM=10 nm			
APPROVAL	WCHENG	JUL./1st/24	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			UVFS	1.5 g	2:1	A