



NOTES

1. MATERIAL: UVFS
2. CLEAR APERTURE: 22 mm
3. CENTER APERTURE: 546 nm
4. FWHM: 11 nm
5. TRANSMITTED WAVEFRONT:  $\lambda @ 633$  nm
6. SURFACE QUALITY: 60/40(S/D)
7. COATING: BANDPASS FILTER, 0° AOI  
 Tabs  $\geq 90\%$  @ 546 nm;  
 ODabs > 5, ODavg > 6 @ 300 nm–532 nm & 559 nm–1000 nm;

DRAWING PROJECTION			<b>LBTEK</b>			
	NAME	DATE	ET546/10x			
DRAWN	LZHOU	JUL./1st/24	PREMIUM BANDPASS FILTER $\phi 25$ mm, CWL=546 nm, FWHM=11 nm			
APPROVAL	WCHENG	JUL./1st/24	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			UVFS	1.5 g	2:1	A