



NOTES

1. MATERIAL: UVFS
2. CLEAR APERTURE: 22 mm
3. CENTER APERTURE: 545 nm
4. FWHM: 24 nm
5. TRANSMITTED WAVEFRONT: $\lambda @ 633 \text{ nm}$
6. SURFACE QUALITY: 60/40(S/D)
7. COATING: BANDPASS FILTER, 0° AOI
 Tabs $\geq 90\% @ 545 \text{ nm}$;
 ODabs ≥ 5 , ODavg $\geq 6 @ 300 \text{ nm} - 526 \text{ nm} \& 563.5 \text{ nm} - 1000 \text{ nm}$;

DRAWING PROJECTION			LBTEK			
	NAME	DATE	ET545/25x			
DRAWN	LZHOU	JUL./1st/24	PREMIUM BANDPASS FILTER $\phi 25 \text{ mm}$, CWL=545 nm, FWHM=24 nm			
APPROVAL	WCHENG	JUL./1st/24	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			UVFS	1.5 g	2:1	A