



NOTES

1. MATERIAL: UVFS
2. CLEAR APERTURE: 22 mm
3. CENTER APERTURE: 266 nm
4. FWHM: 3 nm
5. TRANSMITTED WAVEFRONT:  $\lambda @ 633 \text{ nm}$
6. SURFACE QUALITY: 60/40(S/D)
7. COATING: BANDPASS FILTER,  $0^\circ \text{ AOI}$   
 Tabs >55% @ 266 nm;  
 ODabs >5 @ 255 nm-262 nm;  
 ODabs >5 @ 270 nm-295 nm;

DRAWING PROJECTION			<b>LBTEK</b>			
	NAME	DATE				
DRAWN	LZHOU	JUL ./1st/24	RAMAN FILTER $\phi 25 \text{ mm}$ , CWL=266 nm, FWHM=3 nm			
APPROVAL	WCHENG	JUL ./1st/24	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			UVFS	1.5 g	2:1	A