



NOTES

1. MATERIAL: UVFS
2. CLEAR APERTURE: 22 mm
3. CENTER APERTURE: 785 nm
4. FWHM: 3 nm
5. TRANSMITTED WAVEFRONT: $\lambda @ 633$ nm
6. SURFACE QUALITY(S1, S2): 60/40(S/D)
7. COATING(S1): BANDPASS FILTER, 0° AOI
 Tabs>90%@785 nm
 ODabs>OD5@600-775 nm, 790-1000 nm

DRAWING PROJECTION			LBTEK			
	NAME	DATE				
DRAWN	JLIU	AUG./5st/25	RAMAN FILTER $\phi 25$ mm, CWL=785 nm, FWHM=3 nm			
APPROVAL	WCHENG	AUG./5st/25	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			UVFS	1.8 g	2:1	A