



NOTES

1. MATERIAL: UVFS
2. CLEAR APERTURE: 22 mm
3. CENTER APERTURE: 575 nm
4. FWHM: 50.5 nm
5. TRANSMITTED WAVEFRONT: λ @633 nm
6. SURFACE QUALITY: 60/40(S/D)
7. COATING: BANDPASS FILTER, 0° AOI
 Tabs $\geq 95\%$ @575 nm,
 ODabs ≥ 5 , ODavg ≥ 6 @300 nm–539 nm & 615 nm–700 nm,
 ODabs ≥ 3 @700 nm–790 nm,

DRAWING PROJECTION			LBTEK			
	NAME	DATE	ET575/50m			
DRAWN	LZHOU	JUL./2nd/24	PREMIUM BANDPASS FILTER $\varnothing 25$ mm, CWL=575 nm, FWHM=50.5 nm			
APPROVAL	WCHENG	JUL./2nd/24	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			UVFS	1.5 g	2:1	A