



NOTES

1. MATERIAL: UVFS
2. CLEAR APERTURE: 22 mm
3. CENTER APERTURE: 784.5 nm
4. FWHM: 6 nm
5. TRANSMITTED WAVEFRONT: $\lambda @ 633$ nm
6. SURFACE QUALITY: 60/40(S/D)
7. COATING: BANDPASS FILTER, 0° AOI
 Tabs > 93% @ 785 nm;
 ODabs > 5, ODavg > 6 @ 680 nm–776.5 nm & 792.5 nm–925 nm;
 ODavg > 1 @ 920 nm–1200 nm;
 ODabs > 2 @ 1030 nm–1050 nm;

DRAWING PROJECTION			LBTEK			
	NAME	DATE	RET785/6x			
DRAWN	LZHOU	JUL./1st/24	RAMAN FILTER Ø 25 mm, CWL=784.5 nm, FWHM=6 nm			
APPROVAL	WCHENG	JUL./1st/24	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			UVFS	1.5 g	2:1	A