



NOTE

1. OUTPUT BEAM DIAMETER: 2.0 mm
2. FULL ANGLE BEAM DIVERGENCE: $\leq 0.047^\circ$
3. AR COATING: 1100 nm-1650 nm
4. WAVELENGTH USED FOR ALIGNMENT : 1310 nm
5. USE INTERFACES: SMA

DRAWING PROJECTION			LBTEK			
	NAME	DATE	AFC1310-2.0-SMA			
DRAWN	BSHU	Aug. /26th/24	$\lambda = 1310 \text{ nm}$ SMA BEAM DIAMETER: 2.0 mm			
APPROVAL	WCHENG	Aug. /26th/24	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			D-ZK3	13.5 g	1.5:1	B