



NOTE

1. OUTPUT BEAM DIAMETER  $1/e^2$ : 3.4 mm
2. FULL ANGLE BEAM DIVERGENCE:  $\leq 0.031^\circ$
3. AR COATING:  $R \leq 0.5\%$  @ 1050–1700 nm
4. WAVELENGTH USED FOR ALIGNMENT : 1310 nm
5. USE INTERFACES: FC/APC

DRAWING PROJECTION			<b>LBTEK</b>			
	NAME	DATE				
DRAWN	BSHU	DEC./20th/24	Stationary aspherical collimator			
APPROVAL	WCHENG	DEC./20th/24	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			D-ZK3	10 g	3:1	A