



NOTE

1. OUTPUT BEAM DIAMETER: 2.1 mm
2. FULL ANGLE BEAM DIVERGENCE:  $\leq 0.055^\circ$
3. AR COATING: 1100 nm–1650 nm
4. WAVELENGTH USED FOR ALIGNMENT : 1550 nm
5. USE INTERFACES: SMA

DRAWING PROJECTION			<b>LBTEK</b>			
	NAME	DATE	AFC1550-2.1-SMA			
DRAWN	BSHU	Aug. /26th/24	$\lambda = 1550$ nm FC/APC BEAM DIAMETER: 2.1 mm			
APPROVAL	WCHENG	Aug. /26th/24	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			D-ZK3	13.5 g	1.5:1	B