



NOTE

1. OUTPUT BEAM DIAMETER $1/e^2$: 3 mm
2. FULL ANGLE BEAM DIVERGENCE: $\leq 0.038^\circ$
3. AR COATING: $R \leq 0.5\%$ @ 1050 nm–1700 nm
4. WAVELENGTH USED FOR ALIGNMENT : 1550 nm
5. USE INTERFACES: FC/PC

DRAWING PROJECTION			LBTEK			
	NAME	DATE				
DRAWN	BSHU	DEC./20th/24	Stationary aspherical collimator			
APPROVAL	WCHENG	DEC./20th/24	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			D-ZK3	10g	3:1	A