



NOTE

1. OUTPUT BEAM DIAMETER: 2.2 mm
2. FULL ANGLE BEAM DIVERGENCE: $\leq 0.03^\circ$
3. AR COATING: $R \leq 0.5\% @ 700\text{ nm} - 1100\text{ nm}$
4. WAVELENGTH USED FOR ALIGNMENT : 780 nm
5. USE INTERFACES: FC/PC

| | | | | | | |
|--|--------|--------------|----------------------------------|--------|-------|-----|
| DRAWING PROJECTION | | | LBTEK | | | |
| | NAME | DATE | | | | |
| DRAWN | BSHU | DEC./20th/24 | Stationary aspherical collimator | | | |
| APPROVAL | WCHENG | DEC./20th/24 | MATERIAL | WEIGHT | SCALE | REV |
| FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES | | | D-ZK3 | 9g | 3:1 | A |