



NOTE

1. OUTPUT BEAM DIAMETER $1/e^2$: 4 mm
2. FULL ANGLE BEAM DIVERGENCE: $\leq 0.021^\circ$
3. AR COATING: $R \leq 0.5\%$ @ 600 nm–1050 nm
4. WAVELENGTH USED FOR ALIGNMENT : 780 nm
5. USE INTERFACES: FC/PC

DRAWING PROJECTION			LBTEK			
	NAME	DATE	AFC780-4-PC			
DRAWN	BSHU	DEC./20th/24	Stationary aspherical collimator			
APPROVAL	WCHENG	DEC./20th/24	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			D-ZK3	10 g	3:1	A