



MOUNTING DOUBLE-BORE  
ACCEPTS  $\phi 25.4\text{mm}$  [1in] MINIMUM 3.0mm [0.12in] THICK OPTICS  
6.5mm [0.26in] DEEP  
 $\phi 23.5\text{mm}$  [ $\phi 0.93\text{in}$ ] CLEAR APERTURE

M6X0.25 ADJUSTER SCREWS  
PROVIDES  $\pm 4^\circ$  TOTAL ANGULAR DISPLACEMENT  
( $0.4^\circ$  PER REVOLUTION)  
2 PLACES  
2.0mm [5/64in] HEX

DRAWING PROJECTION			<b>LBTEK</b>			
	NAME	DATE				
DRAWN	CZH	AUG./9th/23	MIRROR MOUNT WITH UNTHREADED HOLE FOR $\phi 25.4\text{mm}$ OPTICS			
APPROVAL	ZXU	AUG./9th/23	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			N/A	0.074 Kg	1:1	C