



NOTES:

1. DESIGN WAVELENGTH: 587.6 nm
2. OPERATION WAVELENGTH: 1050-1700 nm
3. DIAMETER TOLERANCE: $+0.0/-0.1$ mm
4. THKKNES TOLERANCE: ± 0.1 mm
5. FOCAL LENGTH: $f=50.2$ mm $\pm 1\%$
6. BACK FOCAL LENGTH(REF): $bf=48.3$ mm
7. CLEAR APERTURE: $>90\%$ CA
8. SURFACE QUALITY(S/D)(S1,S2): 40/20
9. SURFACE FLATNESS(S1): $\lambda/2 @ 632.8$ nm
10. SURFACE POWER(S2): $3\lambda/2 @ 632.8$ nm
11. SURFACE RREGULARITY(S2): $\lambda/4 @ 632.8$ nm
12. CENTRATION: <3 arcmin
13. CHAMFER: $<0.2\text{mm}, 45^\circ$
14. AR COATNG(S1,S2): $R_{avg}<0.5\% @ 1050-1700$ nm (6° AOI, single)

DRAWING PROJECTION			LBTEK			
	NAME	DATE				
DRAWN	AGAO	MAY./22th/25	PLANO CONVEX LENS			
APPROVAL	SHAN	MAY./22th/25	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			UVFS	0.63g	5:1	A