



#### NOTES

1. DESIGN WAVELENGTH: 587.6 nm
2. OPERATION WAVELENGTH: 700 nm–1100 nm
3. DIAMETER TOLERANCE:  $+0.0/-0.1\text{ mm}$
4. THICKNESS TOLERANCE:  $\pm 0.1\text{ mm}$
5. FOCAL LENGTH:  $f=40.1\text{ mm} \pm 1\%$
6. BACK FOCAL LENGTH(REF):  $bf=35.3\text{ mm}$
7. CLEAR APERTURE:  $>90\%CA$
8. SURFACE QUALITY(S/D)(S1, S2): 40/20
9. SURFACE FLATNESS(S1):  $\lambda/2 @ 632.8\text{ nm}$
10. SURFACE POWER(S2):  $3\lambda/2 @ 632.8\text{ nm}$
11. SURFACE IRREGULARITY(S2):  $\lambda/4 @ 632.8\text{ nm}$
12. CENTRATION:  $<3\text{ arcmin}$
13. CHAMFER:  $<0.2\text{ mm}, 45^\circ$
14. COATING(S1, S2): AR COATING  $R_{avg} < 0.5\% @ 700\text{ nm} - 1100\text{ nm}$

DRAWING PROJECTION			<b>LBTEK</b>			
	NAME	DATE				
DRAWN	BSHU	JUL./24th/24	$\phi 25.4\text{ mm}, f=40.1\text{ mm}$ PLANO CONVEX LENS AR COATING 700–1100 nm			
APPROVAL	WCHENG	JUL./24th/24	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			UVFS	5.78 g	2:1	A