



NOTES

1. OPERATION WAVELENGTH: 1100 nm–1700 nm
2. OPTICAL DENSITY: $0.5 \pm 5\%$ @1300 nm
3. CLEAR APERTURE: >90%CA
4. SURFACE QUALITY(S1,S2): 40/20(S/D)
5. TRANSMITTED WAVEFRONT ERROR: $\lambda/4$ @632.8 nm
6. PARALLELISM(S1,S2): <10 arcsec
7. CHAMFER: <0.2 mm, 45°
8. COATING(S1,S2): AR COATING,
Ravg<0.5%@1100 nm–1700 nm, 6° AOI
9. OTHER SURFACES: FINE GROUND

DRAWING PROJECTION			LBTEK			
	NAME	DATE				
DRAWN	JLIU	OCT./28th/25	ABSORPTIVE NEUTRAL DENSITY FILTER Ø 12.7 mm, OD=0.5, AR: 1100 nm–1700 nm			
APPROVAL	WCHENG	OCT./28th/25	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			ND Filter Glass	0.92 g	4:1	A