



SPLITTING COATING(BEVEL FACE IN THE TRIANGULAR PRISM WITH DOT)

NOTES

1. CLEAR APERTURE: >80%CA
2. AR COATING(S1,S2,S3,S4): $R_{avg} < 0.5\%$ @ 900 nm-1300 nm(6° AOI,SINGLE SURFACE)
3. DIAMETER TOLERANCE: +0.0/-0.25 mm
4. ANGLE TOLERANCE: $\pm 2'$
5. COATING(S5): $T_p, avg > 90\%$, $R_s, avg > 95\%$ @ 900 nm-1300 nm, AOI=45°
6. GLUED(S5,S6): NOA61
7. EXTINCTION RATIO(T_p, T_s): >1000:1
8. TRANSMITTED BEAM DEVIATION: $0^\circ \pm 5$ arcmin
9. REFLECTED BEAM DEVIATION: $90^\circ \pm 5$ arcmin
10. SURFACE QUALITY(S1,S2,S3,S4): 40/20(S/D)
11. TRANSMISSION WAVEFRONT ERROR: $\lambda/2$ @ 633 nm
12. SURFACE FLATNESS(S1,S2,S3,S4): $\lambda/4$ @ 633 nm
13. OTHER SURFACES: FINE GROUND
14. CHAMFER: <0.5 mm, 45°



RATIO 1:2

DRAWING PROJECTION			LBTEK			
	NAME	DATE				
DRAWN	LTAN	Aug./31st/25	POLARIZATION BEAM SPLITTER L=50.8 mm, $T_p, T_s > 1000:1$ @ 900 nm-1300 nm			
APPROVAL	WCHENG	Aug./31st/25	MATERIAL	WEIGHT	SCALE	REV
FOR INFORMATION ONLY NOT FOR MANUFACTURING PURPOSES			N-SF1	401.1 g	1:1	B